

17.02.2017

Amendment No. II

AMENDMENT TO E-TENDER FOR DATA LOGGER – AT IVC


Ref: Tender No: HBL/IVC/LTE/DATA_LOGGER/16-17 dated 31.01.2017

With reference to the Tender Enquiry Document for Data Logger at IVC, the following amendment may please be noted

Data sheet point number and excerpt*/description of the specifications*	Point modified as/ comment
Data Sheet point	Modified as
3.2. Screen – 5.5 inch LCD display	Screen – 12 inch LCD display
3.4. Number of analog input – 30 Channels	Number of input – 30 channels (Provision should be available for 100 channels)
9.3. Calibration certificate should be provided	9.3. Test certificate should be provided
New addition at specific equipment requirement	Thermocouples – 50 Nos T type T/c Sensor for autoclave T type T/c Sensor for DHS

All other terms & conditions, specifications, drawing etc shall remain the same as per the tender No: HBL/IVC/LTE/DATA_LOGGER/16-17 dated 31.01.2017.

For HLL Biotech Ltd.


17/2/17

Authorized Signatory